

资质



国家电工仪器仪表质量监督检验中心
National Quality Supervision and Inspection Center for Electrical Measuring Instruments

CERTIFICATE

TYPE TEST OF ENERGY METERS

Applicant: YUE QING HEYI ELECTRICAL CO.,LTD.
Contact person: Mr Hu Wei
Address: Shangyanhou village Liushi Yueqing Zhejiang China
Postal code, Place: Zhejiang, China 325600
Country: China

Manufacturer: YUE QING HEYI ELECTRICAL CO.,LTD.
Address: Shangyanhou village Liushi Yueqing Zhejiang China
Postal code, Place: Zhejiang, China 325600
Country: China
Electrical apparatus: SINGLE PHASE TWO WITE ELECTRONIC METER
Type: DDS450CTPT
Register: ANALOG/LCD
Class: Active 1.0S
Measurement range: 5(60A) 240V 60HZ 1.0

The energy meter meets the requirements of:

IEC 62053-21:2003 Electricity metering equipment(a.c) Static meters for active energy (classer 1 and 2)

IEC 62052-11-2003 Electricity metering equipment (a.c) General requirements, tests and test cond itions - Part 11: Metering equipment

The results are recorded in our test report with reference WJ14241



Signature: 

Date: March 17, 2015

地址: 中国哈尔滨市松北区创新一路 2599 号 邮编: 150081 电话: 0451-86681896(86611012)
Address: No.2599, Chuangxinyi Road, Songbei District, Harbin, Heilongjiang, China Postcode:150081
Tel: +86-451-86681896(86611012)

Mail: zjzx@tc104.org

Web:www.sicem.org

Fax: +86-451-86681896(86611012)-808